


HR-100 積分球を上から見た図




HR-100 の透過測定

従来法と補償法の測定手順

ゼロ調整と迷光補正の説明は省略


従来法

ASTM D1003-61, JIS K 7105-1981




全光線透過率のみ補償法

ISO 13468-1:1996, JIS K 7361-1:1999



ヘーズのみ補償法

ISO 14782-1:1999, JIS K 7136:2000



すべて補償法

ISO 13468-1:1996, JIS K 7361-1:1999

ISO 14782-1:1999, JIS K 7136:2000

